


Measurement Results

Repeatability

Test Load: 100 g

	As Found	As Left
1	N/A	100.000 g
2	N/A	100.000 g
3	N/A	100.000 g
4	N/A	99.999 g
5	N/A	99.999 g
6	N/A	100.000 g

○ As Found
 ◆ As Left



Standard Deviation	N/A	0.0005 g
--------------------	-----	----------

The "d" in the graph represents the readability of the range/interval in which the test was performed.


The results of this graph are based upon the absolute values of the differences from the mean value.

Eccentricity

Test Load: 100 g

Position	As Found	As Left
1	N/A	0.000 g
2	N/A	0.000 g
3	N/A	0.000 g
4	N/A	0.000 g
5	N/A	0.000 g

Maximum Deviation	N/A	0.000 g
-------------------	-----	---------



As Left

The "d" in the graph represents the readability of the range/interval in which the test was performed.

Error of Indication